

# CERTIFICATE

## of biocompatibility

### Wirobond® 280

Wirobond® 280 is a cobalt-based dental casting alloy. It is suitable for the fabrication of crowns, bridges as well as metal-ceramic restorations.

#### Composition in % by mass:

| Co   | Cr   | W   | Mo  | Ga  | Mn    | Si    |
|------|------|-----|-----|-----|-------|-------|
| 60.2 | 25.0 | 6.2 | 4.8 | 2.9 | < 1.0 | < 1.0 |

The alloy corresponds to ISO 22674 and ISO 9693-1.

Wirobond® 280 is free of nickel, cadmium, beryllium and lead in accordance with ISO 22674.

#### Manufacturer:

**BEGO Bremer Goldschlägerei Wilh. Herbst GmbH & Co. KG**

Wilhelm-Herbst-Str. 1 · 28359 Bremen, Germany

#### Production:

Wirobond® 280 is produced in accordance with ISO 9001 and ISO 13485 and is approved as a class IIa medical device in the EU.

#### Mechanical properties:

The mechanical requirements in accordance with ISO 22674 have been satisfied.

#### Tests and results:

##### Corrosion

A corrosion resistance test in accordance with the ISO 22674 "Metallic materials for fixed and removable restorations and appliances" standard has been performed. The value recorded was considerably lower than the threshold value for the ion release (corrosion) of 200 µg/cm<sup>2</sup> in 7 days defined by the standard. The corrosion resistance and intra-oral biocompatibility are thus confirmed.

##### Cytotoxicity

A test for potential cytotoxicity in accordance with the internationally applicable ISO 10993-5 standard has been performed. No cytotoxic potential was determined.

#### Final evaluation:

It is hereby confirmed that the material has been evaluated in accordance with the internationally applicable EN ISO 10993-1: "Biological evaluation of medical devices" standard. The evaluation includes, amongst other things, possible risks such as cytotoxicity, sensitisation, irritation and genotoxicity.

The tests conducted were performed in independent testing facilities in accordance with the specifications of the OECD guidelines and in compliance with the GLP (Good Laboratory Practice) requirements.

The evaluation confirms the biological compatibility of Wirobond® 280 when used in accordance with the intended purpose.

#### Date of issue:

March 31<sup>st</sup>, 2020



**Dr.-Ing. Stephan Kim**

Chief Development and Innovation Officer

BEGO Bremer Goldschlägerei  
Wilhelm Herbst GmbH & Co. KG



**Dr. Roland Strietzel**

Medical Devices Safety Officer

BEGO Bremer Goldschlägerei  
Wilhelm Herbst GmbH & Co. KG