CERTIFICATE

of biocompatibility

Mediloy® M-Ti5

Mediloy[®] M-Ti5 is a titanium-based dental alloy for milling. It is suitable for the fabrication of crowns and bridges or metal-ceramic crowns and bridges, as well as implant prosthetics.

Composition in % by mass:	Ti	AI	V		
	90,0	6,0	4,0		
	The alloy corresponds to ISO 22674 and ISO 9693-1. Mediloy® M-Ti5 is free of nickel, beryllium, cadmium and lead in accordance with ISO 22674.				
Manufacturer:	BEGO Bremer Goldschlägerei Wilh. Herbst GmbH & Co. KG Wilhelm-Herbst-Str. 1 · 28359 Bremen, Germany				
Production:	Mediloy [®] M-Ti5 is produced in accordance with ISO 9001 and ISO 13485 and is approved as a class IIb medical device in the EU.				
Mechanical properties:	The mechanical requirements in accordance with ISO 22674 have been satisfied.				
Tests and results:	A corrosio restoration than the t	Corrosion A corrosion resistance test in accordance with the ISO 22674 "Metallic materials for fixed and removable restorations and appliances" standard has been performed. The value recorded was considerably lower than the threshold value for the ion release (corrosion) of 200 µg/cm ² in 7 days defined by the standard. The corrosion resistance and intra-oral biocompatibility are thus confirmed.			
	Cytotoxicity A test for potential cytotoxicity in accordance with the internationally applicable ISO 10993-5 standard has been performed. No cytotoxic potential was determined.				
Final evaluation:	It is hereby confirmed that the material has been evaluated in accordance with the internationally applicable EN ISO 10993-1: "Biological evaluation of medical devices" standard. The evaluation includes, amongst other things, possible risks such as cytotoxity, sensitisation, irritation and genotoxicity.				
	The tests conducted were performed in independent testing facilities in accordance with the specifications of the OECD guidelines and in compliance with the GLP (Good Laboratory Practice) requirements.				
	The evaluation confirms the biological compatibility of Mediloy [®] M-Ti5 when used in accordance with the intended purpose.				
Date of issue:	March 31s	^t , 2020			
		pa		('Idand Stretzed	
	-	tephan Kim elopment and	Innovation Officer	Dr. Roland Strietzel Medical Devices Safety Officer	

BEGO Bremer Goldschlägerei Wilhelm Herbst GmbH & Co. KG BEGO Bremer Goldschlägerei Wilhelm Herbst GmbH & Co. KG